

DBH
TLN

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 7/8/2014

Instrument Serial Number: 010460

As of 08-Jul-14

Start Date	End Date	Agency Name
08-Jul-14		DFS Central Lab
27-Oct-10	08-Jul-14	Newport News PD
25-Oct-10	27-Oct-10	DFS Central Lab
11-Dec-09	25-Oct-10	Newport News PD
09-Dec-09	11-Dec-09	DFS Central Lab
09-Dec-09	09-Dec-09	Newport News PD
13-Oct-09	09-Dec-09	DFS Central Lab
09-Oct-09	13-Oct-09	Newport News PD
28-Jul-09	09-Oct-09	DFS Central Lab
07-Jan-09	28-Jul-09	Caroline County VSP Area 44
15-Aug-08	07-Jan-09	DFS Central Lab

DBH
TLN**INTOX EC/IR II
Quality Assurance Worksheet**

Instrument Serial Number 010460 Worksheet Start Date 7/8/2014
Location Newport News Pentran
Address 150 35th St Newport News VA 23607
DFS Technician Donald Hall License No. 21250

☐ Laboratory☒ On-SiteSite Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) 753 Reference Barometer (mm HG) 756
Reference Barometer(RB)Serial # 009109 RB Calibration Due 8/1/2014

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.299
0.300	0.297	0.288	0.305	Sample 2	0.299
Precision		sample min	sample max	Sample 3	0.298
0.001		0.298	0.299		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.099
0.100	0.099	0.096	0.102	Sample 2	0.099
Precision		sample min	sample max	Sample 3	0.099
0		0.099	0.099		

Dry gas standard Lot No. (with tank no.)

AG316301-24

☐ Replaced dry gas standard (+O-ring)☐ Installed at Location☒ Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:

Instrument Serial Number

010460

Certification Date

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section


DFS Technician



Date

7/9/14


Issuing Analyst


MSK

Date

7/9/14
7/10/14
TKW

DBH
TLN
7/10/14




COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, B		AGENCY DFS Central Lab
DFS LICENSE NUMBER 21250	LICENSE EXPIRES 10/01/2014	DATE TEST CONDUCTED 07/08/2014
TEST EQUIPMENT NUMBER 010460		

RESULTS: TIME SAMPLE TAKEN 15:57 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20 _____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____
OPERATOR'S SIGNATURE

DBH
TLN



IntoxNet MIS Report

Report Generated 09 Jul 2014 at 00:24

Test Results

Instrument Serial Number 010460

Test # 002039 Subject Test

Test Location 1 Department of
Test Date 08 Jul 2014

Test Location 2 Forensic Science
Test Time 15:50
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B

Card Serial Number 121250

Effective Date 10/01/2012

License Number 21250

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 08 Jul 2014 End Time 15:58

Result Time 15:57

Result Date 08 Jul 2014 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 15:51

Data Type BLK

Sample Value 0.000

Sample Time 15:51

Data Type CHK

Sample Value 0.099

Sample Time 15:52

Data Type BLK

Sample Value 0.000

Sample Time 15:53

Data Type SUBJ

Sample Value 0.000

Sample Time 15:54

Data Type BLK

Sample Value 0.000

Sample Time 15:55

Data Type SUBJ

Sample Value 0.000

Sample Time 15:57

Data Type BLK

Sample Value 0.000

Sample Time 15:58

Standard Type Dry Gas Std

Standard Value 0.099

Standard Lot Number AG316301-24

Standard Expiration Date 06/12/2015

Tank Pressure 316

Barometric Pressure 753 mmHg

Blow Sample Number 1 Blow Duration 3.48 sec

Blow Volume 1960 cc End-of-Blow Time 15:54

Blow Sample Number 2 Blow Duration 3.17 sec

Blow Volume 1830 cc End-of-Blow Time 15:57

Tamper Evident Stamp e26ada9b

Test Status Code 0

Test Status Success

Intox EC/1R-II: Accuracy Check

DBH
FLV

Department of Forensic Science

Serial Number: 010460 Test Number: 2040

Test Date: 07/08/2014 Test Time: 15:59 EDT

Dry Gas Target: 0.099

Lot Number: AG316301-24 Exp Date: 06/12/2015

Tank Pressure: 311 psi Barometric Pressure: 753 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	16:00
CHK	0.099	16:00
BLK	0.000	16:02
CHK	0.099	16:03
BLK	0.000	16:04
CHK	0.099	16:05

Test Status: *Success*

Calibration CRC: 4C36F00B

Intox EC/IR-II: Accuracy Check

DAJ
TLN

Department of Forensic Science

Serial Number: 010460 Test Number: 2041

Test Date: 07/08/2014 Test Time: 16:07 EDT

Dry Gas Target: 0.297

Lot Number: AG334603-07 Exp Date: 12/12/2015

Tank Pressure: 75 psi Barometric Pressure: 753 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	16:08
CHK	0.299	16:09
BLK	0.000	16:11
CHK	0.299	16:11
BLK	0.000	16:13
CHK	0.298	16:14

Test Status: *Success*

Calibration CRC: 4C36F00B